

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAR 08 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BCC

Agency: HILLSBOROUGH CO SO
Time of Inspection: 20:16

Date of Inspection: 03/03/2017

Serial Number: 80-003388
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.080	0.208	/ 0.084
0.000	0.049	0.081	0.203	/ 0.085
0.000	0.049	0.080	0.202	/ 0.088
0.000	0.050	0.081	0.201	/ 0.088
0.000	0.050	0.081	0.202	/ 0.087
0.000	0.049	0.081	0.201	/ 0.087
0.000	0.049	0.081	0.201	/ 0.086
0.000	0.050	0.082	0.201	/ 0.086
0.000	0.049	0.081	0.202	/ 0.086
0.000	0.049	0.082	0.200	/ 0.086
Standard Deviations	0.0006	0.0006	0.0022	/ 0.0012

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0011 Number of Simulators Used: 5

Remarks:

08: Ambient Fail CLEARED AREA, Control Outside ToleranceUNK. UNKNOWN CONTROL OUTSIDE TOLERANCE, WILL CONDUCT FURTHER TESNon-compliance: UNKNOWN COT, FURTHER TESTING.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

SP BK

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

03/03/2017
Date